

HB 832 -- INSPECTION OF X-RAY SYSTEMS (Brown (57))

COMMITTEE OF ORIGIN: Standing Committee on Health and Mental Health Policy

This bill prohibits cone beam computed tomography systems and panoramic x-ray systems from being required to be inspected more frequently than every six years. The bill defines "cone beam computed tomography system" as a medical imaging device using x-ray computed tomography to capture data using a cone-shaped x-ray beam. The bill defines "panoramic x-ray system" as an imaging device that captures the entire mouth in a single, two-dimensional image including the teeth, upper and lower jaws, and surrounding structures and tissues.

The provisions of the bill will expire December 31, two years after the effective date.